



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/520,981	08/29/2005	Lionel Vedrine	P-5753	5214
7590 02/12/2007 David W Hightet Becton Dickinson and Company 1 Becton Drive-MC 110 Franklin Lakes, NJ 07417			EXAMINER MEHTA, BHISMA	
			ART UNIT 3767	PAPER NUMBER
SHORTENED STATUTORY PERIOD OF RESPONSE		MAIL DATE	DELIVERY MODE	
3 MONTHS		02/12/2007	PAPER	

Please find below and/or attached an Office communication concerning this application or proceeding.

If NO period for reply is specified above, the maximum statutory period will apply and will expire 6 MONTHS from the mailing date of this communication.

Office Action Summary

Application No.

10/520,981

Applicant(s)

VEDRINE, LIONEL

Examiner

Bhisma Mehta

Art Unit

3767

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 11 January 2005.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-8 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-5 is/are rejected.
- 7) ☒ Claim(s) 6-8 is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☒ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 11 January 2005 is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☒ All b) ☐ Some * c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. _____.
 - ☒ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- ☒ Notice of References Cited (PTO-892)
- ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- ☒ Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date 04/04/2005.
- ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____.
- ☐ Notice of Informal Patent Application
- ☐ Other: _____.

DETAILED ACTION

Information Disclosure Statement

1. Part of the information disclosure statement filed April 4 2005 fails to comply with 37 CFR 1.98(a)(2), which requires a legible copy of each cited foreign patent document; each non-patent literature publication or that portion which caused it to be listed; and all other information or that portion which caused it to be listed. A copy of the foreign patent document, FR 2 750 051 A, has not been received.

Priority

2. It is noted that this application appears to claim subject matter disclosed in prior Application No. PCT/FR03/02252, filed July 16 2003. A reference to the prior application must be inserted as the first sentence(s) of the specification of this application or in an application data sheet (37 CFR 1.76), if applicant intends to rely on the filing date of the prior application under 35 U.S.C. 119(e), 120, 121, or 365(c). See 37 CFR 1.78(a). For benefit claims under 35 U.S.C. 120, 121, or 365(c), the reference must include the relationship (i.e., continuation, divisional, or continuation-in-part) of all nonprovisional applications. If the application is a utility or plant application filed under 35 U.S.C. 111(a) on or after November 29, 2000, the specific reference to the prior application must be submitted during the pendency of the application and within the later of four months from the actual filing date of the application or sixteen months from the filing date of the prior application. If the application is a utility or plant application

Art Unit: 3767

which entered the national stage from an international application filed on or after November 29, 2000, after compliance with 35 U.S.C. 371, the specific reference must be submitted during the pendency of the application and within the later of four months from the date on which the national stage commenced under 35 U.S.C. 371(b) or (f) or sixteen months from the filing date of the prior application. See 37 CFR 1.78(a)(2)(ii) and (a)(5)(ii). This time period is not extendable and a failure to submit the reference required by 35 U.S.C. 119(e) and/or 120, where applicable, within this time period is considered a waiver of any benefit of such prior application(s) under 35 U.S.C. 119(e), 120, 121 and 365(c). A benefit claim filed after the required time period may be accepted if it is accompanied by a grantable petition to accept an unintentionally delayed benefit claim under 35 U.S.C. 119(e), 120, 121 and 365(c). The petition must be accompanied by (1) the reference required by 35 U.S.C. 120 or 119(e) and 37 CFR 1.78(a)(2) or (a)(5) to the prior application (unless previously submitted), (2) a surcharge under 37 CFR 1.17(t), and (3) a statement that the entire delay between the date the claim was due under 37 CFR 1.78(a)(2) or (a)(5) and the date the claim was filed was unintentional. The Director may require additional information where there is a question whether the delay was unintentional. The petition should be addressed to: Mail Stop Petition, Commissioner for Patents, P.O. Box 1450, Alexandria, Virginia 22313-1450.

If the reference to the prior application was previously submitted within the time period set forth in 37 CFR 1.78(a), but not in the first sentence(s) of the specification or an application data sheet (ADS) as required by 37 CFR 1.78(a) (e.g., if the reference was submitted in an oath or declaration or the application transmittal letter), and the

Art Unit: 3767

information concerning the benefit claim was recognized by the Office as shown by its inclusion on the first filing receipt, the petition under 37 CFR 1.78(a) and the surcharge under 37 CFR 1.17(t) are not required. Applicant is still required to submit the reference in compliance with 37 CFR 1.78(a) by filing an amendment to the first sentence(s) of the specification or an ADS. See MPEP § 201.11.

Drawings

3. The drawings are objected to under 37 CFR 1.83(a). The drawings must show every feature of the invention specified in the claims. Therefore, the respective means for operating said means of holding the needle in position and said means of holding the container in position must be shown or the feature(s) canceled from the claim(s). Also, the pierceable zone of the piston must be shown or the feature(s) canceled from the claim(s). No new matter should be entered.

Corrected drawing sheets in compliance with 37 CFR 1.121(d) are required in reply to the Office action to avoid abandonment of the application. Any amended replacement drawing sheet should include all of the figures appearing on the immediate prior version of the sheet, even if only one figure is being amended. The figure or figure number of an amended drawing should not be labeled as "amended." If a drawing figure is to be canceled, the appropriate figure must be removed from the replacement sheet, and where necessary, the remaining figures must be renumbered and appropriate changes made to the brief description of the several views of the drawings for consistency. Additional replacement sheets may be necessary to show the renumbering

Art Unit: 3767

of the remaining figures. Each drawing sheet submitted after the filing date of an application must be labeled in the top margin as either "Replacement Sheet" or "New Sheet" pursuant to 37 CFR 1.121(d). If the changes are not accepted by the examiner, the applicant will be notified and informed of any required corrective action in the next Office action. The objection to the drawings will not be held in abeyance.

4. The drawings are objected to because it is unclear what reference numeral 41 in Figure 2 is meant to represent. It is further noted that many of the reference characters, such as reference characters 21 and 27 in Figure 2 are not clear. Corrected drawing sheets in compliance with 37 CFR 1.121(d) are required in reply to the Office action to avoid abandonment of the application. Any amended replacement drawing sheet should include all of the figures appearing on the immediate prior version of the sheet, even if only one figure is being amended. The figure or figure number of an amended drawing should not be labeled as "amended." If a drawing figure is to be canceled, the appropriate figure must be removed from the replacement sheet, and where necessary, the remaining figures must be renumbered and appropriate changes made to the brief description of the several views of the drawings for consistency. Additional replacement sheets may be necessary to show the renumbering of the remaining figures. Each drawing sheet submitted after the filing date of an application must be labeled in the top margin as either "Replacement Sheet" or "New Sheet" pursuant to 37 CFR 1.121(d). If the changes are not accepted by the examiner, the applicant will be notified and informed of any required corrective action in the next Office action. The objection to the drawings will not be held in abeyance.

Specification

5. The abstract of the disclosure is objected to because it contains legal phraseology. Correction is required. See MPEP § 608.01(b).
6. The disclosure is objected to because of the following informalities: reference character 15 is used for a circular ring in line 33 of page 4 and for a distal rib in line 19 of page 5. Appropriate correction is required.

Claim Objections

7. Claims 1-5 are objected to because of the following informalities: Claim 1 recites the limitations "said means for holding the needle in position" and "said means of holding the container in position" in lines 26-27. There is insufficient antecedent basis for these limitations in the claim. Appropriate correction is required.
8. Claims 6-8 are objected to under 37 CFR 1.75(c) as being in improper form because a multiple dependent claim cannot depend from any another multiple dependent claim. See MPEP § 608.01(n). Accordingly, the claims 6-8 have not been further treated on the merits.

Claim Rejections - 35 USC § 102

9. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the

Art Unit: 3767

applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

10. Claims 1-5 are rejected under 35 U.S.C. 102(e) as being anticipated by Barker et al (U.S. Patent No. 6,569,115). Barker et al disclose a device for injecting a product having a body (20) with a hollow injection needle (50) and a container (62) containing the injectable product. The device also has a plunger (30) that slides in the body. The container (62) is closed at one end and is connected to the plunger (30). Means for keeping the needle in position (26, 44) and means for keeping the container in position in the form of the ribs on the forward end of the container (line 59 of column 5 to line 3 of column 6) are provided on the device. Barker et al also disclose the device as including a piston (64) engaged in the container which is so shaped that, in a first configuration, closes the container and, in a second configuration, allows product to pass out of the container. Respective means for operating said means of holding the needle in position and said means of holding the container in position are in the form of the displacement of the needle retainer (40) along with the release of the container (lines 30-49 of column 7). The piston (64) in the second configuration as described in lines 42-58 of column 4 has a pierceable wall or zone which allows the product to pass between the piston and the container. The piston has a peripheral zone (68) which is the section of the piston which presses tightly against the wall of the container, specifically at the rim of the container, as seen in Figure 4, and which is withdrawn from the rim, as seen in Figure 6. Barker et al also disclose the device as having spring means (55).

Art Unit: 3767

Conclusion

11. The prior art made of record and not relied upon is considered pertinent to applicant's disclosure. Botich et al (U.S. Patent No. 6,123,688) disclose a retractable needle injection device.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Bhisma Mehta whose telephone number is 571-272-3383. The examiner can normally be reached on Monday through Friday, 7:30 am to 3:00 pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Kevin Sirmons can be reached on 571-272-4965. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.



BM

